


PO2-15 Percent Oxygen Sensor





Direct Replacement Guide

Analytical Industries: PSR-11-25 Teledyne: B-5

Specifications:

Sensor Technology	Micro Fuel Cell
Signal Output ¹	78 - 122 uA
Measuring Range	0 - 100 Percent
Response Time T90	13 Seconds
Accuracy Full Scale ²	± 1% of Signal
Repeatability F.S.	± 0.5%
Temp Coefficient	2.54 % / ° C
Operating Temp	0 to 50 ° C
Storage Temp	0 to 45 ° C
Recommended Flow	0.5 - 2 SCFH
Humidity	0 - 100% RH (Non-Condensing)
Expected Life ³	34 - 38 Months
Storage	6 Months
Warranty⁴	12 Months
PCB Connection	CenterFoilNegative Outer Foil Positive

1. Signal output is measured in air at 25 ° C, sea level.

2. Accuracy full scale is calculated with constant pressure and temperature & proper calibration (90% O2 value on full scale range). Drastic temperature change can result in a maximum error of \pm 10%.

3. Expected life is calculated when O2 is 20.9% @ 25 $^{\rm o}$ C, sea level.

4. Southland Sensing warrants these sensors for the period noted above to be free from defects in materials and workmanship. Southland Sensing will not be held liable for sensors damaged due to customer neglect.

深圳市新世联科技有限公司